



Hydro-Check
SYSTEMS, INC

Water Quality Instrumentation
Conductivity • %Rejection • TDS • Resistivity

5931 Sea Lion Place
Suite 100
Carlsbad, CA 92010

Tel: 760-930-1924 Fax:
760-930-1934
www.hydrocheck.com

Certificate of Compliance

Model: All HLS & HLA Model Hy-Lites

We hereby certify that all Hy-Lite resistivity indicators are calibrated to a specific resistivity set point at the time of manufacture and are not field adjustable. Final test and calibration of all units is performed in a circulating water system using a Thornton 200CR monitor in current calibration, and hereby certified to be within +/- 3% of specified set point.

Scott Brown, Operations Manager
Scott@Hydrocheck.com