



## Certificate of Compliance

Verification Equipment: Thornton 200CR

Serial Number: 313871

Last Calibration: 3/20/23

Next Calibration: 3/20/25

Model: All 414 Monitor Controllers

This is to certify that all 414 Monitor Controllers were manufactured according to applicable drawings and specifications. They have undergone rigorous testing and demonstrated compliance with established standards. Specifically, it has been verified that all 414 Monitor Controllers are within the calibration accuracy of their respective model values. All 414 Monitor Controllers are calibrated to a specific set point at the time of manufacture and should not be field-adjusted unless necessary. Testing and calibration were conducted in a controlled environment, utilizing a temperature-stabilized, closed, circulating water system and N.I.S.T traceable fixed resistors. The measurements and assessments were performed using the Thornton 200CR instrument, calibrated in accordance with ISO 17025 and ANSI/NCSL Z-540 standards, traceable to the National Institute of Standards and Technology (NIST). This ensures high precision and reliability in the results obtained.

*Matt Rego*

Matt Rego

Production Administrator  
Hydro-Check Systems, Inc

*Adene Hammond*

Adene Hammond  
Vice President  
Hydro-Check Systems, Inc

1708 Ord Way  
Oceanside, CA 92056

760-930-1924  
760-930-1934 Fax